

SETRON pH Meter
Type ARGUS X
Operations manual

Distributor:
Topac Inc
231 CJC Highway
Cohasset
MA 02025
Tel 781 740 8778
Fax 781 740 8779
www.topac.com
[email sales@topac.com](mailto:sales@topac.com)

Warning! There are no serviceable or replaceable parts in this product. Do not remove any covers as this can damage the instrument and compromise warranty.

All information contained in this manual is current at the time of publication. Our commitment to product improvement requires that we reserve the right to change equipment, procedures and specifications at any time.

Manual ARGUS χ , SENTRON Ref. number E7500174, rev. 06, July 2007



In case your pH meter has this label and you are located in Europe, it means that in case the meter cannot be used any more, you have to send it back to Sentron to be destroyed in an environmental safe way. Never put the meter into a trash can for 'normal' waste.

1. INTRODUCTION	5
1.1. WELCOME	5
1.2. DECLARATION OF CONFORMITY	5
1.3. DEFINITION OF THIS MANUAL	5
1.4. LAYOUT OF THE METER	6
1.4.1. Keypad.....	6
1.4.2. Display.....	7
1.4.3. Connections.....	9
1.4.4. Battery	9
2. HELP-FUNCTION.....	10
3. DISPLAY PH AND TEMPERATURE	10
4. CALIBRATION	11
4.1. PERFORMING A CALIBRATION.....	11
4.1.1. 1-point calibration.....	12
4.1.2. 2- or 3- point calibration.....	13
5. AUTOREAD: STABILITY FUNCTION.....	14
6. MEASUREMENT DATA HANDLING	15
6.1. SAMPLE IDENTIFICATION.....	15
6.2. DATA STORAGE	15
6.3. DATA RECALL	16
6.4. DATA EXPORT (PC/PRINTER)	17
6.5. DATA DELETION	18
6.5.1. Delete last recorded data set.....	18
6.5.2. Delete data.....	18
7. ADDITIONAL SYSTEM CONFIGURATION	20
7.1. BIAS OPTION.....	20
7.2. DISPLAY CONTRAST	21
7.3. PARAMETER SETTINGS	21
7.3.1. Select °C/°F.....	21
7.3.2. Power saving options.....	21
7.3.3. Select language.....	22
7.3.4. Calibration buffers	22
7.3.4.1. SENTRON bufferset.....	23
7.3.4.2. DIN bufferset	23
7.3.4.3. NIST bufferset	24
7.3.4.4. JIS bufferset	24
7.3.4.5. Own defined bufferset.....	25
7.3.5. mV reading	27
7.3.6. Calibration data	27
7.3.6.1. Calibration data storage	27
7.3.6.2. Calibration data recall and export	28
7.3.7. Allow measurement data deletion.....	29
7.3.8. Set the “AR” stability criterium.....	30
7.3.9. Calibration interval.....	30
7.3.10. RS-232 configuration.....	30
7.3.11. Date and time	31
7.3.12. Automatic data storage.....	31

- 8. TROUBLESHOOTING33**
 - 8.1. ATT CODES.....33
 - 8.2. RESET36
- 9. AVAILABLE SENTRON PH PROBES37**
- 10. SYSTEM SPECIFICATIONS38**
- 11. DEFAULT SETTINGS39**
- 12. CLEANING PROCEDURE.....40**
- 13. WARRANTY40**
- 14. KEYWORDS.....42**

1. Introduction

1.1. Welcome

Congratulations! You have purchased a SENTRON ARGUS χ pH-meter capable of highly accurate pH measurement by using the reliable and innovative SENTRON ISFET probes.

We advise to read the Quickstart-card carefully and act according to the instructions to ensure that the SENTRON ARGUS χ system will work enjoyably for a long time. This manual can be of help when further details on features are required.

The SENTRON pH-meters and probes are designed for pH-measuring only. Do not use in any other application as this might result in instrument failure or damage.

All wetted materials used are in compliance with the demands of the Food & Drug Administration (FDA) of the United States of America.

Warning! There are no serviceable or replaceable parts in this product. Do not remove any covers as this can damage the instrument and compromise warranty.

1.2. Declaration of Conformity

SENTRON Europe B.V. of Roden, The Netherlands declares that this ARGUS χ system is in compliance with the EMC-norms EN 50081-1 and EN 50082-1.

Due to conformity to these and other norms, the instrument is entitled to wear the CE-mark.

1.3. Definition of this manual

In this manual, words placed between “quotation marks” indicate that this text is shown in the meter’s display. Words between ‘brackets’ indicate that you are prompted to perform an action.

Additional information can be given in a **note** at the end of a section. **Warning!** indicates that potentially harmful actions are to be avoided. **DANGER!** indicates potential hazards when the equipment is improperly used.

At the beginning of this manual a ‘Table of Contents’ gives an overview of its lay-out and indicates where specific information can be found. The last chapter of this manual provides an alphabetic keywords list, referring to the pagenumbers where the information can be found.

In most cases however, the meter’s on-board help function can also be sufficient.

1.4. Layout of the meter

1.4.1. Keypad

The 'controls' of the SENTRON pH meter are condensed to only six keys located on the keypad.

The names 'Power', 'Enter' and 'Arrow' keys ('Left', 'Right', 'Up' and 'Down') will be used consistently throughout this manual.

The backlight can be switched on by simultaneously pressing the 'Left' and 'Up' key. It will switch off automatically after some time (see paragraph 7.3.2).




Figure 1: Keypad ARGUS χ

After starting the meter by pressing 'Power', the display will briefly show the text 'SENTRON integrated sensor technology', the meter type (ARGUS χ), the software revision number and date, and the meter's serial number. Then the Main menu as described hereafter will be shown.

In the bottom left corner of the display an indication can read "ATT001" or another number. This is quite normal and no cause for concern.

Warning ! As the system is not yet calibrated, the pH value shown in the display is not reliable.

1.4.2. Display

The ARGUS X's display is a graphics Liquid Crystal Display (LCD). The pH value can be shown in 1 or 2 decimal resolution and temperature can be shown or be omitted. In chapter 3 these settings are explained in detail.

Paragraph 0 describes how the display contrast can be set over a wide range, allowing easy reading under virtually any ambient light conditions.

In addition, the display can be illuminated by a built-in backlight, that allows reading even in darker environments. The backlight can be switched on by simultaneously pressing the 'Left' and 'Up' key. It will switch off automatically after some time (see paragraph 7.3.2).

Figure 2 shows all possible contents of the display.




Figure 2: Possible display contents

The display is divided in three sections: the top-section is one line, presenting date and time information (see also paragraph 7.3.11).

The large middle section of the display shows following information:

Probe status indicator

Three blocks means that the probe gives maximum performance.

Two blocks means that the probe functions fine. Some maintenance (cleaning procedure, see page 40 of this manual) can bring it up to maximum performance. In case two blocks are achieved by new probes, it is advised to place the probe-tip in hand-warm water for 20 minutes, and then in buffer pH 4 for 1-2 hours. Recalibrate.

One block is typical for a probe that has been in use for some time. The probe still gives accurate results, but may require cleaning by using water, toothbrush and some mild detergent. Recalibrate using fresh buffers.

If none of the recommended remedial actions lead to probe status improvement, the probe is near to the end of its functional life. Replacement will be necessary in the near future.

pH 12.34 **pH value indication**

These large digits represent the pH value

- 123.4 °C **Temperature indication**

These smaller digits represent the actual temperature in °C/°F, see paragraph 7.3.1.



Battery status indication

When the remaining operational time is less than approximately 5 hours, a “low” indication will appear. When charging the battery, two alternating “+” symbols are shown on the right hand side of the battery symbol.


The filling of the battery symbol indicates the actual battery status.

AR **Autoread activated**

When the Autoread function (automatic stability check) is activated, the text “AR” will be shown in the display. Both the pH value and the “AR” symbol will blink when the signal is not stable. When the pH measurement signal is within the stability limits, the pH value and the “AR” symbol will be shown continuously, thus providing a clear visual stability indication. Also refer to chapter 5.

CAL **Calibration indication**

In the bottom left corner the text “CAL” will be blinking when calibration is in progress.



Calibration prompt

When the programmed time between calibrations has elapsed, the “CAL” text and this clock-symbol will be displayed.

Status indication text line

The status text line at the bottom of the middle section is used for a wide variety of indications. Examples can be found throughout this manual.

Help Disp Calib
AutoR Store Config **Lower section menu items**

In the lower section, the operating menu allows a variety of choices. Use the ‘Arrow’ keys to go to the desired selection, then use the ‘Enter’-key to activate the highlighted function. The menu options shown here are from the Main menu. Every sub-menu will show its own menu items.

1.4.3. Connections

Physically different connections guarantee that it is impossible to fit connectors to the wrong receptacle.

The probe needs to be connected at the handle of the ARGUS χ meter (not shown). After insertion the connector needs to be screwed on hand-tight by turning the sleeve clockwise.

Figure 3 shows the backside of the cradle where an RS-232 connector and an AC Power inlet are provided.




Figure 3: Connectors available on the cradle of the ARGUS χ .

From left to right, following connectors are available:

RS-232 connection. Through a standardized 9-pole sub-D connector data can be sent to an external device. Paragraph 7.3.10 gives detailed information on the serial communication possibilities.



AC Power supply. SETRON provides adapters that transform wall-outlet AC power to the level required by the ARGUS χ meter. The meter itself takes care of charging the internal battery. See paragraph 1.4.4 for more information on battery management.



Note To optimize battery life the external outputs (relay outputs, mV output and RS-232 data output) are only active when the meter is AC powered.

1.4.4. Battery

The built-in battery is a rechargeable Nickel-Metal-Hydride (NiMH) battery. Before shipment from SETRON the battery is fully charged, but we suggest to connect to AC power for at least 12 hours before relying only on battery power. Battery charging is indicated in the display by a “++” sign on the right hand side of the battery symbol.

A fully charged battery will typically give 24 hours of use. The battery status is constantly indicated in the display, and automatic ‘low’ battery warning is given when the remaining battery life is approximately 5 hours.

Should the battery be completely exhausted during normal operation, it is still possible to recharge the battery to its normal condition but it is advised that the battery status indication is monitored by the user and that the battery is timely recharged by connecting the meter to AC power.

Note Overcharging the battery is not possible and the applied NiMH-batteries do not suffer from the so called “memory-effect”.

Continued on next page!

A prolonged exhausted battery may result in an automatic meter reset. After recharging a completely exhausted battery, a user-reset may be required to restart the meter. Please refer to paragraph 8.2.

First connect the meter to the mains, and only then switch it off. This allows a visual check if the battery is charging correctly, indicated by the “++” signs next to the battery symbol.

2. HELP-Function

The ARGUS χ meter is equipped with an elaborate on-board help function, designed to provide to-the-point assistance.

The HELP function presents a situation dependant overview of the meter's functionality which will in most cases be sufficient for the user to continue operation.

If a situation occurs that requires the user to take actions (indicated by the text “ATTxxx” on the display) the helptext will accurately describe the situation and what remedy can be taken.

3. Display pH and Temperature

The ARGUS χ meter is a highly sophisticated and accurate instrument, equipped with a high quality graphic Liquid Crystal Display (LCD).

The pH value can be shown with either one- or two-decimal resolution, depending on the calibration method, the user's preference and the application demands.

The sample's temperature is measured by a thermistor which is built into the tip of the probe and provides fully Automatic Temperature Compensation (ATC). The temperature can be indicated in the display.

Thus, the system offers four possible display settings:

- pH 0.1 without temperature indication
- pH 0.1 with temperature indication
- pH 0.01 without temperature indication
- pH 0.01 with temperature indication

To change between these settings, go to the Main menu. Use the 'Arrow' keys to go to “Disp” and press 'Enter' repeatedly until the desired setting appears.

Note Because of the nature of a 1-point calibration, the accuracy is not sufficient for a 0.01 pH indication. When selecting the 0.01 pH resolution indication after a 1-point calibration, the second decimal will blink to indicate that it is not to be regarded as accurate. It is therefore advised to perform a calibration that provides the accuracy required by the application, i.e. a 1-point calibration for 0.1 pH accuracy or a 2- or 3-point calibration for 0.01 pH accuracy.

4. Calibration

To obtain reliable readings from the ARGUS χ system it must first be calibrated, using correct buffers, for maximum accuracy at the same temperature as the sample will be.

The ARGUS χ has four built-in tables of 5 buffers versus temperature and a possibility to enter an own-defined bufferset. When performing a calibration, the bufferset chosen is shown on the display. The built-in buffersets are standardized DIN, JIS and NIST buffers and a SENTRON provided NIST traceable bufferset.

Factory default, the SENTRON bufferset is selected. Please refer to paragraph 7.3.4 on selection of another bufferset.

The system can be set to remind the user to re-calibrate after a certain period of time. Paragraph 7.3.9 provides all details on this function.

The system automatically stores calibration results for later review. Paragraph 7.3.6 provides all details on this function.

Warning ! Make sure that the buffers used during calibration are identical to the selected bufferset, otherwise significant variations in measurement values may occur.

Over time, the value from a buffer may change. Especially buffers with values over pH 7.00 are susceptible for CO₂ contamination.

Make sure that the buffers used for calibration are fresh, and not contaminated by other materials.

SENTRON provides buffers in twin-neck bottles specially designed to facilitate this use.

4.1. Performing a calibration

If the probe is newly connected to the meter (indicated by ATT021 on the status text line), or if the meter has been switched off for several hours, it is necessary to place the probe in a buffer solution, with the meter switched on, for a period of 10 minutes. This initiating time allows the probe to set itself for use after a period of inactivity and is required to ensure stable readings.

Rinse the probe before calibration.

The ARGUS χ allows for various calibration methods. In general, the achievable measurement accuracy over a certain pH range will increase when a more-point calibration is performed.

During calibration, the "CAL" indication will blink in the display

After completing the calibration, the slope percentage is briefly displayed, except with the 1-point calibration as no slope can be calculated there.

Instead, the last calculated slope is used.

Slope values can be interpreted as follows:

Slope %	Interpretation
98.0 – 105.0 %	Probe gives maximum performance
94.0 – 97.9 %	<p>Probe functions fine. Some maintenance, cleaning with water, toothbrush, and a mild detergent as described in the Probe Insert can bring it up to maximum performance.</p> <p>In case such low slopes are achieved by new probes, it is advised to place the probe-tip in hand-warm water for 20 minutes, and then in buffer pH 4.00 for 1-2 hours. This will effectively remove any KCl crystallization that may have occurred in the reference diafragma due to long storage.</p>
90.0 – 93.9 %	<p>Typical for a probe that has been in use for some time. Probe still gives accurate results, but may require cleaning.</p> <p>Use water, toothbrush and a mild detergent as described in the Probe Insert to ensure optimum performance, then recalibrate. If the slope remains low, use fresh buffers and recalibrate. If the slope is close to 90% and no longer improves by mentioned remedies, the probe may be aging and a replacement should be purchased.</p>

Following sections describe the various procedures to perform a calibration.

Note For maximum accuracy, choose buffers that are close to, or bracket, the expected sample pH and perform the calibration around the same temperature as the expected sample temperature.

If a calibration has not been successful, a message to that effect will be shown on the display. Also, the user may opt to cancel the calibration procedure before it is finished. In both cases, the meter will use the data from the last successful calibration to work with.

4.1.1. 1-point calibration

The measurement accuracy achieved is ± 0.1 pH and when the display is set to two-decimal resolution the second decimal will blink continuously to indicate that it is not to be regarded as accurate.

In the Main menu, go to “Calib” and press ‘Enter’.

On the status text line the selected bufferset is shown.

Put the probe in the buffer solution.

Go to “1 Pnt” and press ‘Enter’.

The text on the status text line changes to “First buffer”, and after that to “Recognized XX.XX”. The menu item “Cal 1” is highlighted.

Check if the value of the buffer recognized by the meter is the same as the buffer in use. If not, go to “Set” and press ‘Enter’ repeatedly until the correct buffer value is shown on the status text line, then return to “Cal 1”.

With “Cal 1” highlighted, press ‘Enter’ to confirm the selection.

The text on the status text line will change to “Stabilizing” and the pH value is blinking until a stable signal is reached.

The 1-point calibration is now complete, and the meter automatically returns to the Main menu.

4.1.2. 2- or 3- point calibration

In the Main menu, go to “Calib” and press ‘Enter’. On the status indication line the selected bufferset is shown. Put the probe in the first buffer solution.

Go to the selected calibration method (i.e. “2 Pnt” or “3 Pnt”) and press ‘Enter’.

The text on the status text line changes to “First buffer” and after that to “Recognized XX.XX”. The menu item “Cal 1” is highlighted.

Check if the value of the buffer recognized by the meter is the same as the buffer in use. If not, go to “Set” and press ‘Enter’ repeatedly until the correct buffer value is shown on the status text line, then return to “Cal 1”.

With “Cal 1” highlighted, press ‘Enter’ to confirm the selection.

The text on the status text line changes to “Stabilizing” and the pH value is blinking until a stable signal is reached. Then the status text line changes to “Next buffer” and after some time to “Recognized XX.XX”. The text in the menu item reads “Cal 2”. Rinse the probe and place it in the next buffer solution.

Check if the value of the buffer recognized by the meter is the same as the buffer in use. If not, use the ‘Arrow’ keys to go to “Set” and press ‘Enter’ repeatedly until the correct buffer value is shown on the status text line, then return to “Cal 2”

With “Cal 2” highlighted, press ‘Enter’ to confirm the selection.

The text on the status text line will change to “Stabilizing” and the pH value is blinking until a stable signal is reached.

Repeat this sequence for subsequent calibration points.

After the last calibration point, the slope(s) will briefly be displayed and the meter automatically returns to the Main menu.

A 2-point calibration will give one calculated slope, a 3-point calibration will give 2 calculated slopes. The slope data is automatically recorded for later use as described in paragraph 7.3.6.

5. Autoread: Stability Function

The Autoread function provides a fully automatic stability check which shows the user when a measurement value is stable.

In the Main menu, go to “AutoR” and press ‘Enter’.

The text “Autoread is ON” will briefly be shown in the status text line.

The “AR” symbol is permanently shown in the display (see section 1.4.2).

Pressing ‘Enter’ once more disables the Autoread function. The text “Autoread is OFF” will briefly show in the status text line, and the “AR” symbol is no longer shown in the display.

When the Autoread function is active, both the “AR” indication and the pH value will blink until the stability criterion is met, providing an easy visual indication of when the measurement has become stable.

The stability criterium is default set at 0.02 pH per 30 seconds.

The ARGUS χ allows the user to modify this criterium as described in paragraph 7.3.8.

Note Even though the stability criterium is set at a 30-second interval, the system is able to determine stability within 8 seconds by using an advanced extrapolation algorithm.

Even when a measurement value is immediately stable, the system still needs 8 seconds to confirm stability; Hence, the measurement value will blink at least 8 seconds.

SENTRON systems continuously monitor the sample’s pH value, i.e. the stable value is not fixed on the display, and when the sample’s pH changes, the value shown on the display will change accordingly.

6. Measurement data handling

The ARGUS χ can store 300 Measurement Data Sets, including date and time.

These Measurement Data Sets can be recalled on the display for future reference, and be exported to an external device through the serial RS-232 output.

6.1. Sample identification

In order to simplify data recognition, a 9-character identification is allocated to the sample. The first six characters are alphanumeric, the last three characters are numeric only.

In the Main menu, go to “Store” and press ‘Enter’.

The status text line in the display shows the text “xxx free XXXXXX-xxx”

“xxx free” indicates the number of free data memory slots (see paragraph 6.5).

The “XXXXXX-xxx” (default: SAMPLE-000) text is the sample identification and can be changed.

Go to “Name&#” and press ‘Enter’

The first letter of the sample identification is highlighted, and can be changed to any alphanumeric character (including ‘space’) by pressing the ‘Up’ or ‘Down’ key.

Press the ‘Right’ key to go to the next character. Use above procedure to set the complete name. The name will be constant with each stored Measurement Data Set.

Moving further to the right, set the three-digit sequential number to the desired start value by pressing the ‘Up’ or ‘Down’ key. When storing a Measurement Data Set, this number will automatically increase by 1.

Press ‘Enter’ to return to the Store menu. The system can now be used to store data.

6.2. Data storage

The actual data stored depends on the chosen display setting, i.e. if the temperature is not shown on the display, it will not be stored either. If the pH value is shown in one-decimal accuracy or if the last calibration was a 1-point calibration, the pH value stored will have a one-decimal accuracy.

Data can be stored either manually or fully automatically.

For a description of automatic data storage, please refer to paragraph 7.3.12.

In the Main menu, go to “Store” and press ‘Enter’.

The status text line in the display shows the text “xxx free XXXXXX-xxx”.

Go to “Meas” and press the ‘Enter’ key to save the data in the meter’s memory.

The “xxx free” indication in the status text line will decrease by one, and the 3-digit sample identification number will increase by one.

The sample identification can be changed at any time by following the procedure described in paragraph 6.1.

Note When the meter’s memory is full (300 Measurement Data Sets stored) a warning will be displayed. The meter will not overwrite previously stored data.

6.3. Data recall

Measurement Data Sets stored in the meter’s memory can be recalled onto the display. The same format will be used as when the data were saved.

In the Main menu, go to “Store” and press ‘Enter’.

Go to “Recal” and press ‘Enter’.

The last measurement stored is shown on the display as follows:




Figure 4: Recall menu

The probe status symbol is not shown in this menu. The text “Recall” is blinking on the status text line. All other data reflect the Measurement Data Set stored in the meter’s memory, i.e. the sample identification, pH value, temperature, date and time.

The highlighted menu item is “Prev”. By pressing ‘Enter’, the previous Measurement Data Set is shown on the display. By going to “Next” and pressing ‘Enter’ the next Measurement Data Set in the meter’s memory is shown on the display.

The other two menu items “Selct” and “All” can be used to export or delete data from the meter’s memory.

Go to “Back” and press ‘Enter’ to return to the Main menu.

6.4. Data export (PC/Printer)

Data stored in the meter's memory can be exported to an external device, e.g. a computer or printer, by using the serial RS-232 output.

To configure the RS-232 port, please refer to paragraph 7.3.10.

To export data, go to "Store" and press 'Enter'.

Go to "Recal" and press 'Enter'.

Data to be exported must be selected first.

To select all data in the meter's memory, use the 'Arrow' keys to go to "All" and press 'Enter'.

To select individual Measurement Data Sets, go to "Selct" and press 'Enter' to select the Measurement Data Set shown on the display.

Go to "Prev" or "Next" and press 'Enter' to show other Measurement Data Sets on the display. Again, go to "Selct" and press 'Enter' to select the Measurement Data Set shown on the display.

Selected Measurement Data Sets are marked with an asterisk (*) in front of the sample identification.

When the Measurement Data Set selection is complete, go to "More" and press 'Enter'. The highlighted menu item is "Send".

Pressing 'Enter' exports the selected Measurement Data Sets to the RS-232 output.

The RS-232 connection to the external device must be programmed correctly to allow correct datatransfer (see paragraph 7.3.10).

The data will be sent in a format as shown in figure 5:

===== STORED DATA =====

1999 Aug 02 10:20:05

SENTRON pH Measurement System

Type: ARGUS χ

S/N: 9300

Last calibration:

Date	Time	Status	Type
99Aug02	08:31	3	SENTRON
Buffer	Temp	Slope	
4.00	18.0	*****	
7.00	18.1	99.8	

Measurement points:

Date	Time	Samplename	pH	Temp A
99Aug02	08:33:06	CHEESE000	4.98	12.0
99Aug02	08:34:06	CHEESE001	4.95	11.8
99Aug02	08:35:06	CHEESE002	7.98	12.0 H

Report generated by:

===== END =====

Figure 6: Stored data report

The start and end of the report are clearly indicated. The date and time of the transmission, the type of meter and the meter's serialnumber are shown in the top of the report, as well as the complete data from the last calibration.

Below that, the selected measurement points are shown in the same format as when they were stored.

Note Both "Selct" and "All" are toggle functions. Thus, it is possible to first select some Measurement Data Sets using "Selct" and then go to "All".
By pressing 'Enter', the previously selected Measurement Data Sets are unselected and all others are selected. This function can be used to export a large number of Measurement Data Sets with the exception of a few.

6.5. Data deletion

To prevent accidental data deletion, the possibility to delete data can be disabled. Paragraph 7.3.7 gives details on this feature.

Regardless of this setting, it is always possible to delete the last recorded data set.

6.5.1. Delete last recorded data set

To facilitate deletion of an accidentally stored dataset, the menu item "AutoM" will change to "Dellst" when a dataset is stored.

To delete the last stored data set, go to "Dellst" and press 'Enter'.

The menu item will change back to "AutoM" and it is not possible to delete further data sets.

The amount of free memory, indicated by "xxx free" will increase by 1 and the sample identification will be adjusted as well.

6.5.2. Delete data

The ARGUS χ has a memory that allows storage of 300 Measurement Data Sets. In the Store menu, the amount of free memory slots is continuously indicated in the status text line as "xxx free".

To delete data, the data to be deleted must be selected first.

In the Main menu, go to "Store" and press 'Enter'.

Go to "Recal" and press 'Enter'.

To select all data in the meter's memory, go to "All" and press 'Enter'.

To select individual Measurement Data Sets, use the 'Arrow' keys to go to "Selct" and press 'Enter' to select the Measurement Data Set shown on the display.

Use the 'Arrow' keys to go to "Prev" or "Next" and press 'Enter' to show other Measurement Data Sets on the display. Again, use the 'Arrow' keys to go to "Selct" and press 'Enter' to select the Measurement Data Set shown on the display.

Selected Measurement Data Sets are marked by an asterix (*) in front of the sample identification.

When the Measurement Data Set selection is complete, go to "More" and press 'Enter'. Go to "Clear" and press 'Enter'.

The text "Cleared sel. Samples" is briefly shown on the status text line, and all selected data are deleted from the memory permanently.

Note Both “Selct” and “All” are toggle functions. Thus, it is possible to first select some Measurement Data Sets using “Selct” and then go to “All”.
By pressing ‘Enter’, the previously selected Measurement Data Sets are unselected, and all others are selected. This function can be used to delete a large number of Measurement Data Sets with the exception of a few.

Warning ! Deletion of data is a non-reversible action. It may be advisable to send the data to an external device prior to deletion (see paragraph 6.4)

7. Additional system configuration

In this chapter a variety of options for additional system configuration are discussed.

7.1. Bias option

In some cases it is possible that there is a difference in measurement results between a Sentron pH measuring system and a classic glass electrode system. This difference can be eliminated by entering a "bias".

Use the arrow keys to go to the menu option "bias". This is a submenu of option "config". The default settings for the bias option is zero. Use the arrow keys to alter the bias value. The pH measurement result will be corrected with the entered bias value. During calibration, the bias is not used.

2001 Dec 31 23:59:59

▣■■■■ ++

pH 12.34

- 123.4 °C

Help	Disn	--
AutoR	Store	Config

2001 Dec 31 23:59:59

▣■■■■ ++

pH 12.34

- 123.4 °C

new bias value: xx.xx

Help		--
LCD	Params	Back

7.2. Display contrast

Specific ambient light situations may require a different display contrast.

To adjust the display contrast setting, in the Main menu, go to “Config” and press ‘Enter’. Go to “LCD” and press ‘Enter’.

The status text line will show “Contrast: x”.

Use the ‘Up’ or ‘Down’ key to change the display contrast setting.

When the contrast is optimal, press ‘Enter’ to confirm the selection and return to the Configuration menu.

Go to “Back” and press ‘Enter’ to return to the Main menu.

Warning ! It is possible to change the display contrast to a setting where it is very difficult to read the display at all. Reverse the display setting immediately to a readable contrast.

7.3. Parameter settings

7.3.1. Select °C/°F

In the Main menu, go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Select degree C/F” and press ‘Enter’.

Press the ‘Up’- or ‘Down’ arrow to change between °C and °F, then press ‘Enter’ to confirm the selection.

7.3.2. Power saving options

Especially with battery powered meter operation, energy conservation is important to optimize the operational time of the meter.

The ARGUS χ can automatically switch itself off after a pre-defined period of inactivity.

In the Main menu, go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Power saving options” and press ‘Enter’.

The Power saving menu is displayed as follows:

Power saving menu

Back

Help

Meter

Display

Backlight

ISFET

Figure 7: Power saving menu

Go to “Meter” and press ‘Enter’.

The text: “Stay ON time of meter (min): xx” is displayed. By using the ‘Up’ or ‘Down’ key, the ON time can be set from 1 minute to 55 minutes or Continuous.

Press ‘Enter’ to return to the Power saving menu.

Use the ‘Down’ key to go to “Display” and repeat the above procedure.

The ON time for the backlight can be set between 5 and 100 seconds in a similar manner.

The ON time for the ISFET control circuit can be set from 1 to 23 hours in a similar manner.

Go to “Back” and press ‘Enter’ to return to the Main menu.

Note As long as the ISFET control circuit is active, the ISFET is powered and remains ‘active’. When the probe is placed in a liquid during this time, it is not necessary to go through the 10-minute start-up procedure.

As the backlight consumes relatively much energy it is advised to limit the ON time when working on battery power.

7.3.3. *Select language*

The meter can be set to one of four languages:

- English
- German
- French
- Spanish

To change the language, in the Main menu go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Select language” and press ‘Enter’.

Use the ‘Down’ or ‘Up’ key to go to the desired language and press ‘Enter’.

Go to “Back” and press ‘Enter’ to return to the Main menu.

7.3.4. *Calibration buffers*

Following sections list the buffersets built into the meter as Buffervalue versus Temperature. The SENTRON Bufferset is the factory set default.

To change to another bufferset, in the Main menu go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Calibration buffers” and press ‘Enter’.

Go to the desired bufferset and press ‘Enter’.

A message “Selected: BUFFERSETNAME” will appear to confirm the selection.

Also, when entering the Calibration menu, the bufferset used will be shown on the status text line.

Go to “Back” and press ‘Enter’ to return to the configuration menu.

Once more, the selected bufferset and its nominal values at 25 °C are shown on the display.

Go to “Back” and press ‘Enter’ to return to the Main menu.

7.3.4.1. *SENTRON bufferset*

The SENTRON bufferset is the factory default setting. It is a set of 5 NIST-traceable, SENTRON provided buffers with pH values 2.00, 4.00, 7.00, 10.00 and 12.00 at 25 °C. The full table of Buffervalue versus Temperature is given below.

°C	#1	#2	#3	#4	#5
5	2.01	4.00	7.10	10.22	12.70
15	2.01	4.00	7.04	10.14	12.38
25	2.00	4.00	7.00	10.00	12.00
35	2.00	4.02	6.98	9.91	11.63
45	2.00	4.04	6.96	9.81	11.38
55	2.00	4.07	6.96	9.76	11.12
65	2.00	4.10	6.95	9.71	10.86
75	2.00	4.12	6.94	9.66	10.60
85	2.00	4.15	6.93	9.61	10.34
95	2.00	4.17	6.92	9.56	10.08

Table 1: SENTRON Bufferset

Note These values are commonly used and are readily available from other suppliers as well.

7.3.4.2. *DIN bufferset*

Below table shows the Buffervalue versus Temperature of the DIN buffers stored in the ARGUS χ memory:

°C	#1	#2	#3	#4	#5
5	3.12	4.67	6.87	9.43	13.57
15	3.08	4.65	6.82	9.32	13.17
25	3.06	4.65	6.79	9.23	12.75
35	3.04	4.65	6.77	9.14	12.45
45	3.04	4.67	6.76	9.05	12.14
55	3.04	4.69	6.76	8.96	11.83
65	3.03	4.71	6.75	8.87	11.52
75	3.02	4.73	6.75	8.78	11.21
85	3.01	4.75	6.74	8.69	10.90
95	3.00	4.77	6.74	8.60	10.59

Table 2: DIN Bufferset

7.3.4.3. *NIST bufferset*

Below table shows the Buffervalue versus Temperature of the NIST buffers stored in the ARGUS χ memory:

°C	#1	#2	#3	#4	#5
5	1.67	4.00	6.95	9.40	13.21
15	1.67	4.00	6.90	9.28	12.81
25	1.68	4.01	6.87	9.18	12.45
35	1.69	4.02	6.84	9.10	12.13
45	1.70	4.05	6.83	9.04	11.84
55	1.72	4.08	6.83	8.99	11.57
65	1.73	4.11	6.84	8.94	11.33
75	1.75	4.15	6.85	8.90	11.11
85	1.78	4.18	6.87	8.88	10.91
95	1.81	4.23	6.89	8.83	10.73

Table 3: NIST Bufferset

7.3.4.4. *JIS bufferset*

Below table shows the Buffervalue versus Temperature of the JIS buffers stored in the ARGUS χ memory:

°C	#1	#2	#3	#4	#5
5	1.67	4.01	6.95	9.39	13.21
15	1.67	4.00	6.90	9.27	12.81
25	1.68	4.01	6.86	9.18	12.45
35	1.69	4.02	6.84	9.10	12.14
45	1.70	4.04	6.83	9.04	11.84
55	1.72	4.08	6.84	8.99	11.58
65	1.73	4.11	6.84	8.94	11.33
75	1.75	4.14	6.85	8.91	11.11
85	1.78	4.18	6.87	8.87	10.91
95	1.81	4.23	6.89	8.83	10.73

Table 4: JIS Bufferset

7.3.4.5. *Own defined bufferset*

The ARGUS χ allows the user to enter an own defined table of Buffervalue versus Temperature. Factory default, the values are similar to the SENTRON bufferset.

Table 5 on the next page can be used to note down the pH values prior to entering them in the meter’s memory and for future reference.

In the Main menu, go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Calibration buffers” and press ‘Enter’.

Go to “Own defined buffer” and press ‘Enter’.

Go to the first buffer to change and press ‘Enter’.

The display shown is given in below figure 8:

Buffer 1 entry	
Temp (C)	pH
5	x.xx
15	x.xx
25	x.xx
35	x.xx
45	x.xx
55	x.xx
65	x.xx
75	x.xx
85	x.xx
95	x.xx

Help
Back

Figure 8: Entry table for Own defined bufferset

Go to a pH value and press ‘Enter’.

The first decimal of the buffervalue will be highlighted, and can be changed with the ‘Up’ or ‘Down’ key. The temperature can not be changed. Use the ‘Left’ or ‘Right’ key to go to the next digit.

Repeat this until the whole pH value is correctly set, then press ‘Enter’.

Repeat this procedure with all necessary Temperature/pH values.

Go to “Back” and press ‘Enter’.

Go to the next buffer if required.

Repeat this whole procedure for all other buffers that need to be set.


Go to “Back” and press ‘Enter’ to return to the Main menu.

Note The “Own defined bufferset” menu can be left at any moment. Make sure that ALL Temperature/ Buffervalues of interest have been entered correctly. Not entering a pH-value can result in a non-sense value should later be calibrated in that Temperature/pH area.

Buffers have to be entered in the correct order, i.e. #1 is lower pH value than #2, etcetera.

°C	#1	#2	#3	#4	#5
5					
15					
25					
35					
45					
55					
65					
75					
85					
95					

Table 5: Own defined bufferset



°C	#1	#2	#3	#4	#5
5					
15					
25					
35					
45					
55					
65					
75					
85					
95					

Table 5: Own defined bufferset

7.3.5. mV reading

The ISFET mV output can be viewed as well.

In case the meter is not yet calibrated for the connected probe, the actual ISFET mV output will be shown. If the meter has already been calibrated for the connected probe, the mV output related to the value at pH 7.00 is shown.

In the Main menu, go to "Config" and press 'Enter'.

Go to "Params", press 'Enter' again.

Go to "mV reading" and press 'Enter'.

The text "ISFET mVolt: xxx.x" is shown on the display. The mV value reading remains active and only pressing 'Enter' brings back the Configuration menu.

To return to the Main menu, go to "Back" and press 'Enter'.

Note When performing a manual slope calculation, note that for an ISFET a theoretical sensitivity of 52,5 mV/pH should be used.

7.3.6. Calibration data

The ARGUS χ allows for storage, retrieval and export of the last 10 calibration data sets. Following sections describe the various options available.

7.3.6.1. Calibration data storage

When performing a calibration, all calibration data are automatically stored in the meter's memory and can be recalled.

Information stored includes the date and time, bufferset used, buffervalue(s), temperature, slope(s) and the probe status.

Data can be recalled on the screen (see section 7.3.6.2) and be exported using the RS-232 output (see section 7.3.10).

7.3.6.2. Calibration data recall and export

In the Main menu, go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Calibration data”.

The display will show the data from the last calibration in the following way:

2001 Dec 31		23:59:59
<hr/>		
■■■■		
Buffer	Temp	Slope
12.34	123.4°C	*****
12.34	123.4°C	123.4
12.34	123.4°C	123.4
Recall 3 pnt calib		
<hr/>		
Help	Next	Prev
Send 1	SndAll	Back

Figure 9: Calibration data recall

Go to “Prev” and press ‘Enter’ to bring the previous calibration data set on the display. Similarly, “Next” brings the next data set on the display.

The menu items “Send 1” and “SndAll” can be used to export data using the serial RS-232 output. Make sure that the parameters are set to match those of the external device (see paragraph 7.3.10).

“Send 1” sends the dataset that is on the display, and “SndAll” sends all stored datasets to the RS-232 output. Go to either selection and press ‘Enter’ to activate the transmission.

The following type of report will be transmitted:

===== CALIBRATION DATA REPORT =====

1999 Aug 02 10:20:05

SENTRON pH Measurement System

Type: ARGUS χ

S/N: 9300

Date	Time	Status	Type
99Aug02	08:31	3	SENTRON
Buffer	Temp	Slope	
4.00	18.0	*****	
7.00	18.1	99.8	

Date	Time	Status	Type
99Aug01	16:31	2	SENTRON
Buffer	Temp	Slope	
4.00	18.5	*****	
Slope used: 97.9			

Report generated by:

===== END =====

The start and end of the report are clearly indicated. The date and time of the transmission, the type of meter and the meter's serialnumber are shown in the top of the report.

Below that, the calibration data is shown, consisting of the date and time of the calibration, the probe status and the chosen bufferset. The buffervalue and temperature are shown, as well as the calculated slope.

The calibration data are printed in sequential order, beginning with the last calibration.

With a 1-point calibration no slope can be calculated and the last calculated slope is used.

The report above is an example of "SndAll" where 2 calibrations were stored in the meter's memory. A maximum of 10 calibrations can be stored by the meter. In case the "Send 1" command was chosen, only the last calibration at the top of the report would have been shown.

7.3.7. Allow measurement data deletion

In the Main menu, go to "Config" and press 'Enter'.

Go to "Params", press 'Enter' again.

Go to "Clear allowed on/off" and press 'Enter'.

The text "Clear data allowed" is briefly shown on the display, and the user can now delete data from the memory as described in section 6.5.2.

To prevent data deletion, press 'Enter' again. The text "Clear data disabled" is briefly shown on the display and the user will no longer be able to delete data.

7.3.8. Set the "AR" stability criterium

The ARGUS χ allows the user to set the stability criterion over a bandwidth from 0.01 to 0.10 pH per 30 seconds. The factory default value is 0.02 pH per 30 seconds.

In the Main menu, go to "Config" and press 'Enter'.

Go to "Params", press 'Enter' again.

Go to "Autoread option" and press 'Enter'

The text "AR (pH/30sec): 0.xx" is shown on the display.

Use the 'Up' or 'Down' key to set the stability criterion to the desired value.

Press 'Enter' to confirm the selection and return to the Main menu.

7.3.9. Calibration interval

In the Main menu, go to "Config" and press 'Enter'.

Go to "Params", press 'Enter' again.

Go to "Calibration interval" and press 'Enter'

The text "Calib interval (h): x" is shown on the display (factory default is 8 hours).

Use the 'Up' or 'Down' key to set the number of hours after which the user must be reminded to recalibrate the system.

Press 'Enter' to confirm the selection and return to the Main menu.

The calibration reminder will appear on the display as the text "CAL" with a clock symbol.

7.3.10. RS-232 configuration

The ARGUS χ handles data transmission to an external device (e.g. computer or printer) by serial connection, using 8 databits, 1 stopbit, no parity. This is a standard setting that is most commonly used and accepted by other equipment. These parameters are factory set and can not be changed.

The transmission speed is determined by the baudrate and is user-configurable in standard steps between 1200 and 19200 baud. Factory default is 9600 baud

Make sure that the same baudrate is used by the ARGUS χ and the external device, as otherwise a miscommunication will occur.

The ARGUS χ can transmit four different reports:

- Last calibration data report (see section 7.3.6.2)
- Last 10 calibration data reports (see section 7.3.6.2)
- Stored data report (see section 6.4)
- Real-time data transmission (see section 6.2)

On these reports, distinguished columns of data are presented. The separation character between the columns is factory set at SPACE but can be changed by the user to , : ; or TAB. This separation character can be used by external devices (e.g. computer software) to distinguish between different columns of data.

In the Main menu, go to "Config" and press 'Enter'.

Go to "Params", press 'Enter' again.

Go to "Set RS232 port" and press 'Enter'.

The text "RS232 Baud: 9600" is shown on the display.

Use the 'Up' or 'Down' key to change the baudrate to the desired setting.

Pressing 'Enter' confirms the baudrate and changes the text on the display to

“Separator: SPACE”. the ‘Up’ or ‘Down’ key can be used to change the separator character to the desired setting.

Press ‘Enter’ to confirm the separator character, then go to “Back” and press ‘Enter’ to return to the Main menu.

Note The RS-232 connection is only active when the meter is AC powered.

7.3.11. *Date and time*

In the Main menu, go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to ‘Set date and time’ and press ‘Enter’.

The display will show ‘Year: xx’ with the first digit highlighted. Use the ‘Up’ or ‘Down’ key to set the correct number, then use the ‘Right’ or ‘Left’ key to go to the second digit. Again, use the ‘Up’ or ‘Down’ key to set the correct number.

Press ‘Enter’ to display the month (01=January, 12=December).

Repeat this procedure to set the day, hours, minutes and seconds. After having set ‘seconds’ press ‘Enter’ to confirm and return to the Main menu.

7.3.12. *Automatic data storage*

The ARGUS χ allows for automatic data storage, acting on parameters set by the user.

In the Main menu, go to “Config” and press ‘Enter’.

Go to “Params”, press ‘Enter’ again.

Go to “Automeasure options” and press ‘Enter’

Below text is shown on the display:

Automeasure menu

Help

Start time

Time interval

of measurements

Back

Figure 10: Automeasure menu

Go to “Start time” and press ‘Enter’.

Any start time within 24 hours can be selected, but it should be noted that this is the actual start time and not a delay time.

The text “Automeasure start Hours: xx” is shown with the first digit highlighted.

Use the ‘Up’ or ‘Down’ key to set the desired start hour, use the ‘Left’ or ‘Right’ key to change to the second digit.

Press ‘Enter’ to go to “Minutes” and repeat the sequence, press ‘Enter’ again for “Seconds”. Once more pressing ‘Enter’ brings back the “Automeasure menu”.

Go to “Time interval” and press ‘Enter’

The interval time defines the amount of time between two measurements.

The text “Automeas interval Hours: xx” is shown with the first digit highlighted.

Any interval time within 24 hours can be selected. Use the ‘Up’ or ‘Down’ key to set

the desired interval hour, use the 'Left' or 'Right' key to change to the second digit. Press 'Enter' to go to "Minutes" and repeat the sequence, press 'Enter' again for "Seconds". Once more pressing 'Enter' brings back the "Automeasure menu".

Use the 'Up' or 'Down' key to go to "# of measurements" and press 'Enter'

The text "# of meas:xxxx" is shown with the first digit highlighted.

Any amount of measurements upto 9.999 can be selected, but only a maximum of 300 Measurement Data Sets will be stored in the meter's memory. If the total number of measurements exceeds the available memory slots, then a message to that effect is displayed.

Use the 'Up' or 'Down' key to set the desired number, use the 'Left' or 'Right' key to change to the second digit.

When the number of measurements is set, press 'Enter' to confirm the selection and return to the Automeasure menu.

The automeasure parameters are now set.

Go to "Back" and press 'Enter' to return to the Main menu.

Go to "Store" and press 'Enter'.

Go to "AutoM" and press 'Enter' to activate Automeasure.

The text "Automeasure is ON" will briefly be displayed in the status text line.

At the programmed start time, the meter will take a first measurement, and continue to take measurements as defined.

The Automeasure option can be disabled by pressing 'Enter' again.

The text "Automeasure is OFF" will briefly be displayed in the status text line.

To return to the Main menu, use the 'Arrow' keys to go to "Back" and press 'Enter'.

Note When the meter's memory is full (300 Measurement Data Sets stored) a warning will be displayed. The meter will not overwrite previously stored data, but will still output the new data to the RS-232 port (provided that AC power is connected)

It is allowed to use the power saving options (see paragraph 7.3.2) as the meter will automatically re-activate itself whenever a new measurement needs to be taken.

When an auto-measurement is ongoing, the system can still be calibrated.

The meter will skip measurements during the calibration and continue after the calibration is finished.

Warning ! When the meter is recalibrated during measurements, only this last calibration will be shown on the Stored Data Report, even though earlier measurements have been taken with another calibration. In such case, exporting a report for the last 10 calibration data sets and comparing the timestamps with the timestamps of the measurement data sets will give adequate information.

8. Troubleshooting

8.1. ATT Codes

The ARGUS χ incorporates a comprehensive set of diagnostic functions, allowing it to automatically detect a variety of unwanted situations. In such a case, the text “ATTxxx” will be shown on the status text line.

The built-in helptexts will state the situation at hand and recommend a remedial action. In case of multiple messages, the ARGUS χ itself handles priorities to make sure that only one ATT code is shown.

ATT000:	No probe signal	Ensure that the probe is properly connected to the meter. Ensure that there is a good contact between the sample and the probe. Clean the probe according to the instructions in the Probe Insert.
ATT001:	The pH is out of the 0-14 range.	Verify the buffer and the buffer recognition during calibration. The probe may not have been calibrated to the meter; Recalibrate. The probe may be contaminated, clean the probe; Recalibrate.
ATT002:	The sample reading is more than 3 pH units out of the calibrated range.	For the best results, calibration should always be bracketing the expected pH-value. Re-calibrate with a wider calibration range or use more calibration points.
ATT003:	The pH is out of the 0-14 range during calibration.	Check if the probe is properly immersed in the buffer. Clean the probe according to the instructions in the Probe Insert and recalibrate.
ATT004:	Less than 0.5 pH units difference between two or more calibration points.	When the calibration points are too close together, accurate slope calculation is difficult. Check the buffers and use other buffers if necessary. Check the buffer recognition during calibration.
ATT005:	The used calibration points are more than 7 pH-units apart.	When the calibration points are too wide apart, the results are less reliable. Check the buffer recognition during calibration. Recalibrate using fresh buffers. Use more buffers and more calibration points.

ATT006:	The ISFET output is too close to the ISFET output in the previous buffer.	Check if the probe is in the right buffer. Recalibrate using fresh buffers. If this does not help, replace the probe.
ATT007:	The probe signal is not stable.	Allow 10 minutes stabilization before calibration. Make sure there is good contact between probe and buffer. Stir buffer well before reading. Clean the probe according to the instructions in the Probe Insert and recalibrate.
ATT010:	There is no thermistor signal.	Without this signal, temperature compensation (ATC) is not possible. Check the proper connection of the probe to the meter. If this does not help, replace the probe.
ATT011:	The temperature is out of the probe's specified working range.	The probe may get damaged. Check the sample temperature. Check if the proper type of probe is used. Refer to the temperature specification of the probe in use.
ATT020:	There is no signal from the β -resistor.	Internal probe malfunctioning. Check the proper connection of the probe to the meter. Replace the probe.
ATT021:	The meter detects a new probe.	Perform a new calibration with this probe.
ATT050:	The slope is too low (slope less than 90).	The sensitivity of the sensor is too low. Recalibrate using fresh buffers. Check for correct buffer recognition during calibration. Check if the used set of buffers matches the programmed set of buffers. Clean the probe according to the instructions in the Probe Insert and recalibrate.
ATT051:	The slope is too high (slope over 110).	The sensitivity of the sensor is too high. Check for correct buffer recognition during calibration. Check if the used set of buffers matches the programmed set of buffers. Clean the probe according to the instructions in the Probe Insert and recalibrate.

ATT060:	Probe status is too low.	Recalibrate using fresh buffers. Check for correct buffer recognition during calibration. Check if the used set of buffers matches the programmed set of buffers. Clean the probe according to the instructions in the Probe Insert.
General	Meter unresponsive	Check times used for Power Saving. Refer to paragraph 7.3.2 Reset the meter. Refer to paragraph 8.2
	No display	Battery may be completely discharged. Connect meter to AC power and wait 10 minutes before turning on the meter. Refer to section 1.4.4 Contrast setting may be set to extreme value. Reverse contrast setting to readable value. Refer to paragraph 0. If unable to find contrast setting menu, reset the meter. Refer to paragraph 8.2.
	Occurrence of strange pixels	Reset the meter. Refer to paragraph 8.2.
	No data or illegible characters sent to external device	Check RS-232 cable for proper connection. Check that baudrate is the same for ARGUS χ and external device. In case of printer, check that a serial printer is used. Set both devices to 1200 baud.
Probe	Probe status	See description in paragraph 1.4.2 Monitoring the Probe status as shown on the Calibration data report over a period of time can result in a more accurate lifetime prediction.
	Clean probe.	The probe can be thoroughly cleaned by washing the tip in luke-warm water (approx. 40 °C). After remaining contamination has been washed off, use the brush and some mild detergent to brush the sensor surface and reference. Put the probe in a saturated KCl solution and let it cool to room temperature. When the probe is cooled down, calibrate again.
Password	Password lost/forgotten	A MASTER Password may be requested from SENTRON by filling out the MASTER Password Request Form in the back of this manual and sending this by fax to SENTRON. SENTRON regards such information as confidential and proprietary and will only release the MASTER password upon receipt of a completely filled out Password Request Form.

8.2. Reset

In case the user has tuned the display contrast to an unreadable setting, it may be required to reset the meter.

Also, unforeseen circumstances that hamper the meter's functioning may require a meter reset.

Reset the meter by depressing and holding the 'Power' key for 40 seconds. The meter may be seen to switch on and off a number of times.

After releasing the 'Power' key the meter will restart automatically within 5 seconds. It may be necessary to re-install any of the following:

- Display contrast (paragraph 0)
- Date and time (paragraph 7.3.11)
- Automeasure option parameters (paragraph 7.3.12)
- After a reset the battery indicator shows low, loading is recommended.

All other data and parameters are automatically saved and restored by the system.

The system will use the data from the last calibration, even though no data are shown in the "calibration data" menu, and the probe status indicator shows no blocks.

Recalibrating the system is recommended.

Note: A prolonged exhausted battery may result in an automatic meter reset.

It is advised that the battery indication is monitored by the user and that the battery is timely recharged by connecting the meter to AC power.

9. Available SENTRON pH probes

This instrument can only perform up to its specifications in combination with one of the SENTRON ISFET pH probes.

The SENTRON “Hot-line” series of probes is developed for applications where sample temperature levels up to 105 °C can be reached. Probe-tips available in the Hot-Line series are Standard, ConeFET, SurFET, LanceFET and LanceFET with handle.

The SENTRON “Stream-line” series of probes is specially developed for measurement in low-conductivity or highly contaminating applications. The reference liquid reservoir can be refilled with a SENTRON provided refill liquid. “Stream-line” probes can also be used up to 105 °C. Probe-tips available in the Stream-Line series are ConeFET and SurFET.

Typical applications for the Standard ISFET are general applications involving pH measurements in liquids, as well as very small sample volume measurements in case of rare or expensive samples, e.g. in medical or cosmetic environments. By placing a drop of liquid in the curved probe tip, pH measurements on small sample volumes can be performed.

The LanceFET can be used for extremely viscous samples, or samples that are difficult to penetrate. The LanceFET with handle ensures good grip when force must be applied to perform a measurement, or when the measurement must be done in cold environments where gloves are worn. Typical applications where these probes are used are pH measurements in cheese, meat and fish.

The SurFET is especially designed for direct pH measurements on flat surfaces, e.g. on wood, textile or paper.

In case of viscous or sticky samples, the ConeFET is advised as this tip is shaped to ensure that remaining sample can easily be cleaned off. Typical applications involve pH measurements in pasta, dough, jelly, etc.

DANGER ! The LanceFET and LanceFET with handle probes are especially designed for easy penetration. The stainless steel tip is very sharp and can cause bodily harm when not properly shielded.

The line of Micro Electrodes are thin electrodes with a diameter of only 3 or 4 mm. In this line we have the standard model of PEEK (3 mm diameter) and the stainless steel model (4 mm diameter). These electrodes can be used for applications in meat, cheese, but also in very small volumes or micro titer plates.

10. System specifications

Measuring range:	pH	0.00 ... 14.00.
	Temperature:	0 °C ... 60 °C or –5 °C ...105 °C probe dependent.
Resolution:	pH	0.1 or 0.01 pH-unit, user selectable.
	Temperature:	0.1 °C.
Accuracy:	pH	± 0.01 + Least Significant Digit (LSD) using 2 or more point calibration. ± 0.1 using 1 point calibration.
	Temperature:	± 0.5 °C
Display:	Backlighted graphics LCD	
Calibration:	1-, 2- or 3-point calibration with automatic buffer recognition.	
Buffer tables:	Sets of DIN, NIST and JIS buffers, a set of 5 SETRON supplied, NIST traceable buffers. One set of 5 user-defined buffers.	
Data logging:	300 measurement data, incl. date and time can be stored, also user programmable start-, interval- and total number of measurements.	
Output:	Galvanically isolated RS-232 serial output with programmable baud-rate.	
Environment:	Temperature:	Measurement 0 °C to 40 °C storage –5 °C to 70 °C.
	Rel. Humidity:	Measurement 95%, storage 95%.
Power supply:	AC Mains with NiMH-rechargeable battery back-up	
Battery charging time:	24 hours when empty, with non-operating meter	
Dimensions (incl. cradle):	Length:	200 mm
	Height:	100 mm
	Width:	85 mm
	Weight:	550 gram
Note	When operating in high humidity-levels, in case of a sudden drop of temperature, problems may arise due to condensation.	
EMC Immunity	When using the AC adapter some interference may be observed at EN50082-1 limits. If unstable readings are observed which cannot be related to the sample, the condition of the probe or meter; try to obtain stable reading on a different location, by RF shielding the sample or by disconnecting the AC adapter.	

11. Default settings

Following table shows the minimum and maximum values and the factory set default value for various parameters.

Parameter		Minimum	Maximum	Default
Power saving Options:	Meter	1 minute	55 minutes/ Continuous	Continuous
	Display	1 minute	55 minutes/ Continuous	Continuous
	Backlight	5 seconds	100 seconds	15 seconds
	ISFET	0 hour	48 hours	8 hours
Autoread	On/Off			Off
	Stability	0.10 pH/30 sec	0.01 pH/30 sec	0.02 pH/30 sec
Automeasure	On/Off			Off
	Start	00:00:00	23:59:59	12:00:00
	Interval	00:00:10	23:59:59	00:00:30
	# of Meas.	0	300	10
Cal. Interval		0 Hours	48 Hours	8 Hours
RS-232	Baudrate	1200	19200	9600
	Separation char	SPACE , ; TAB		SPACE
	Databits	Factory set		8
	Stopbit			1
	Parity			No
Password	On/Off			Off
	Nr.	0000	9999	0000
Clear allowed	Allowed/disabled			Disabled
Cal. Buffers	SENTRON	See section 7.3.4.1		Default
	DIN	See section 7.3.4.2		
	NIST	See section 7.3.4.3		
	JIS	See section 7.3.4.4		
	Own def.	See section 7.3.4.5		As SENTRON
Display	Contrast	-22	+22	0
	pH Resolution	0.1	0.01	0.1
	Temperature	°C	°F	°C
	Temperature	Not shown	Shown	Shown
Time		00:00:00	23:59:59	C.E.T.
Sample ID	6 Alphanumeric and 3 numeric	AAAAAA-000	ZZZZZZ-999	SAMPLE-000
Language	English, German, French, Spanish			English

Note ISFET 0 hour means that the ISFET control circuit will power down when the meter is shut off, either manually or automatically.

12. Cleaning procedure

In many cases of returned probes, the probe does not work due to a polluted diaphragm. This is caused by the fluid in which the end-user is measuring. If the customer does not clean the probe regularly, the diaphragm will block the internal electrode. In such case there is no electrical contact between the electrode and the Isfet chip and the probe will not work.

To prevent this, the end-user must clean the probe regularly, the frequency is depending on the sample the end-user is measuring in. As reference: it is mostly also visible, the diaphragm is no longer white!

Cleaning:

Put the probe in soapy water with a constant temperature of 60 °C for minimal 5 minutes. After this the probe must be placed directly in a KCL solution or buffer 7 with a temperature of ± 20 °C for 1/2 hour. After these actions you can start cleaning with the soft toothbrush.

After measurements in proteins we advise to put the probe alternate in buffer 10 and buffer 3, rinse with water or soak in NaOH or KCl, before you start the cleaning procedure.

Scratches:

To avoid scratches on the chip/epoxy, it is advised to flush the probe with water and to use a wetted brush only after the cleaning procedure as described above (in soapy-water).

Most scratches occur when there are hard pieces on the ISFET and together with the brush they can occur scratches on the surface of the chip and damage the epoxy around the chip, which give bad readings.

FLUSH BEFORE BRUSH!!

13. Warranty

This SENTRON meter is produced, packed and shipped with the utmost care. If, the former notwithstanding, defects do arise, be advised that this SENTRON meter is warranted to be free from defects in material and craftsmanship for the period of 12 months.

SENTRON will repair or replace, at SENTRON's option, any defective part free of charge if this product fails within 24 months from the date of purchase, provided that the failure is due to defective material or lack of craftsmanship and has occurred under normal conditions of usage, to be judged by SENTRON.

All SENTRON probes have a 6 month limited warranty, please refer to the probe manual for specifics.

SENTRON disclaims any liability to customers, to users of its products, or to any other person or persons for any special or consequential damage that might arise out of or that might in any way be connected with the use of this instrument or its accessories.

The warranty described in this paragraph shall be in lieu of any other warranty, expressed or implied, including but not limited to any implied warranty or merchantability or fitness for a particular purpose. The buyer's sole and exclusive remedy is for repair or replacement of defective parts as provided therein.

Representations and warranties made by any person, including dealers, representatives and employees of SENTRON, which are inconsistent or in conflict with the terms of this warranty, shall not be binding upon SENTRON unless in writing and signed by one of its officers.

SENTRON reserves the right to ask for proof of purchase, such as the original invoice or packing slip.

Warning! There are no serviceable or replaceable parts in this product. Do not remove any covers as this can damage the instrument and compromise warranty.

14. Keywords

A

AC Power	9
Accuracy.....	11; 38
Alarm.....	8; 11
AR	8; 14
Arrow-keys.....	6; 8
Assistance	10
ATC.....	10
ATT Code.....	33
Automeasure.....	31
AutoR	14
Autoread	8; 14; 30

B

Backlight	6; 7; 22; 38
Battery	9; 36; 38
Battery charging	9; 38
Battery low	8; 9
Battery status	8
Baudrate	30; 38
Bias option.....	20
Buffers.....	11; 38
Bufferset	11; 22

C

CAL.....	8
Calibration	8; 10; 11; 38
Calibration data	27
Calibration data report.....	28
Calibration interval	30
Calibration prompt.....	8
Calibration, 1-point	10
Celcius	21
CE-mark	5
Cleaning procedure.....	43
Connector	9
Contrast	7; 21

D

Data logging	38
Databit	30
Date and time.....	7; 31
Default value	39
Delete	18
Delete data.....	29
Dellst	18
Dimensions.....	38
DIN Buffers.....	11; 23
Display	7; 10; 21; 38
Down-key.....	6

E

EMC.....	5; 38
Energy conservation.....	21
Enter-key.....	6; 8
Environment.....	38

F

Fahrenheit	21
------------------	----

H

Help function	10
Highlight	8
Hot-line	37
Humidity	38

I

Illuminate	7
Initiate	11

J

JIS Buffers	11; 24
-------------------	--------

K

Keypad	6; 8
--------------	------

L

Language.....	22
LCD	7; 10; 21; 38
Left-key.....	6
LSD.....	38

M

Measurement Data Set	18
Measuring range.....	38
Memory.....	18
Memory-effect	9
Menu item	8
mV reading.....	27

N

NIST Buffers.....	11; 24
-------------------	--------

O

Operating menu 8
 Output 9; 38
 Overcharging 9
 Own defined buffers 25

P

Parity 30
 pH value 7; 8; 10; 38
 Power supply 38
 Power-key 6
 Probe 37

R

Relay output 9
 Reset 36
 Resolution 7; 10; 38
 Right-key 6
 RS-232 9; 15; 17; 28; 30; 38

S

Sample identification 15
 SAMPLE-000 15

Select language 22
 SENTRON Buffers 11; 23
 Serial communication 9; 17
 Serial number 6
 Slope 11
 Slope calculation 27
 Software revision number 6
 Specification 37; 38
 Stability 14
 Stability criterion 14; 30
 Start time 31
 Start-up 11
 Status text line 8
 Stopbit 30
 Stream-line 37

T

Temperature 7; 8; 10; 11; 21; 22; 38
 Time 7; 31

U

Up-key 6

W

Warranty 2; 41